



MGB Endoskopische Geräte GmbH Berlin

Geschäftsführung / CEO
Schwarzschildstraße 6
12489 Berlin – Germany
Tel.: +49 (0)30 6392-7000
Fax: +49 (0)30 6392-7002
E-Mail: info@mgb-berlin.de
www.mgb-berlin.de

MGB · Schwarzschildstraße 6 · 12489 Berlin - Germany

To all customers

Berlin, 19.02.2019

Urgent security information:
Medical product safety information

Reference: R-2019-1

Affected products: GRP Handles 429-77015(N), 429-77020(N), 429-77030(N)

Dear customer/dealer,

With this letter we would like to inform you that the MGB Endoskopische Geräte GmbH Berlin initiates an unsolicited recall of the GRP handles with and without ratchet due to a change of performance.

This recall applies to all products from the period 01/2008 to 01/2019.

- GRP handles with HF-connector REF: 429-77015(N) LOT: AE...AQ1
GRP handles with HF-connector and ratchet REF: 429-77020(N) LOT: AE...AQ1
GRP handles with HF-connector and ratchet REF: 429-77030(N) LOT: AE...AQ1

1 Encoded monthly: 1st letter: month (A: January, B: February,... M: December)
2nd letter: year (E: 2008, F: 2009,... Q: 2019)

This corrective action is carried out with the knowledge of the competent supervisory authorities.

Table with 2 columns: Health risk, Actions. Health risk: Low risk of electric shock and possible burns of the surgeon. Actions: 1. Check your inventory... 2. Send the completed return form... 3. Return the blocked products... 4. The return form must also be completed... 5. Please ensure that this safety information is passed on... 6. To ensure the effectiveness of the action...



MGB Endoskopische Geräte GmbH Berlin

Geschäftsführung / CEO
Schwarzschildstraße 6
12489 Berlin – Germany
Tel.: +49 (0)30 6392-7000
Fax: +49 (0)30 6392-7002
E-Mail: info@mgb-berlin.de
www.mgb-berlin.de

It is the goal of MGB Endoskopische Geräte GmbH Berlin to deliver only the highest quality products and to provide you with unlimited support.

We are actively investigating this matter to find a long-term solution as fast as possible.

Please return all affected handles together with the attached return form to MGB as far as possible until 22.02.2019:

MGB Endoskopische Geräte GmbH Berlin

Schwarzschildstr. 6
12489 Berlin - Germany

Tel.: +49 (0)30 6392-7040

Fax: +49 (0)30 6392-7011

E-Mail: recall@mgb-berlin.de
www.mgb-berlin.de

Please return the return form even if you have no GRP handles from us in your stock.

All returned products will be credited to your customer account.

Availability of alternative products:

We will contact you and introduce you to available alternatives that may be considered substitutes.

We regret any inconvenience that you, your patient or your staff may have incurred or may have incurred in connection with this recall.

If you have any questions or need more information, please contact our service department at the numbers listed above.

We thank you for your support and understanding.

General Management
Dongjoo Kang, Kwang Seog Song
Tel.: +49 (0)30 6392-7000
Fax: +49 (0)30 6392-7002
www.mgb-berlin.de

Handelsregister HRB 61110
Amtsgericht
Berlin-Charlottenburg
Steuernr.: 37/12720116
USt-IdNr.: DE184229983

Landesbank Berlin
Kto.-Nr.: 1523337032
BLZ: 10050000
IBAN: DE92100500001523337032
SWIFT-BIC: BELADEBE

Commerzbank AG
Kto.-Nr.: 0699999200
BLZ: 10080000
IBAN: DE61100800000699999200
SWIFT-BIC: DRESDEFF100



MGB Endoskopische Geräte GmbH Berlin

Geschäftsführung / CEO
Schwarzschildstraße 6
12489 Berlin – Germany
Tel.: +49 (0)30 6392-7000
Fax: +49 (0)30 6392-7002
E-Mail: info@mgb-berlin.de
www.mgb-berlin.de

MGB · Schwarzschildstraße 6 · 12489 Berlin - Germany

MGB Endoskopische Geräte GmbH Berlin
Schwarzschildstraße 6
12459 Berlin

Please send this form via mail to the address on the left. Alternatively you can send us a signed scan per email to recall@mgb-berlin.de or via fax to +49 (0)30 6392 7011. Thank you!

Return form R-2019-1

Hereby we confirm the reception of the security information for the recall:

Reference: R-2019-1

In our stock we have

- affected products we will return (quantity: pieces).
no affected products.

Company:
Distributor:
Name:
Address:
Email:
Phone:
Comment:

Date

Signature

With my signature I assure that I informed all my customers and/or users about the recall and took every necessary actions.