

Siemens Healthcare Diagnostics GmbH, SHS EMEA CEET QT, Siemensstrasse 90,
1210 Vienna

Name M.A. Roland Ertl
 Department SHS EMEA CEET QT
 Telephone +43 51707-38274
 Mobile +43 (664) 8011738274
 E-mail roland.re.ertl@siemens-healthineers.com
 Date July 24th, 2019
 Document Ref# IMC19-07.A.OUS

– **Urgent Field Safety Notice:**

**Immulite® / Immulite® 1000 / Immulite® 2000 / Immulite® 2000 XPi
 Low Discordant Progesterone results on a Subset of Patient Samples**

Dear Sirs,

Our records indicate that your facility may have received the following product:

Table 1. IMMULITE Systems Affected Product(s)

Assay	Test Code	Catalog Number	Siemens Material Number (SMN)	Lot Number	Manufacturing date
PRG	PRG	LKPW1	10381128	Kit lots 0259 and above	2018-10-19
PRG	PRG	L2KPW2 L2KPW6	10381181 10381170	Kit lots 510 and above	2018-10-01 2018-09-28

Reason for Communication

The purpose of this communication is to inform you of a potential issue with the performance of the products indicated in Table 1.

Siemens Healthcare Diagnostics Inc. has determined that there is a potential for low discordant progesterone results on a subset of patient samples. Our investigation of customer complaints suggests the presence of a potential interferent with the assay. Based on available data, it is estimated the occurrence is <1%.

Due to the nature of initial complaints received, the investigation to date has focused on In Vitro Fertilization (IVF) patient sample performance. Siemens is actively working to investigate the root cause.

Risk to Health

A falsely low progesterone result may lead to the consideration for additional progesterone supplementation. Progesterone results would be used in conjunction with the patient's medical history, clinical examination and other findings such as serial hCG measurements and ultrasound. Siemens is not recommending a review of previously generated patient results.

Actions to be Taken by the Customer

- Please review this letter with your Medical Director.
- If interference is suspected, follow your established internal laboratory procedures to investigate the interference.
- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- If you have received any complaints of illness or adverse events associated with the products listed in Table 1, immediately contact your local Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Sincerely yours,

Siemens Healthcare Diagnostics GmbH

A handwritten signature in black ink, appearing to read "Franz Schwarz".

i.V. Dipl. Ing. Franz Schwarz
Quality Management CEE

A handwritten signature in blue ink, appearing to read "Gernot Osterer".

i.A. Ing. Gernot Osterer
Product Manager Austria & SEE