



TRACOE medical GmbH | Reichelsheimer Str. 1 / 3 | 55268 Nieder-Olm

**TRACOE medical GmbH**

Reichelsheimer Straße 1 / 3  
55268 Nieder-Olm / Germany

Phone: +49 6136 9169-0  
Fax: +49 6136 9169-200

info@tracoe.com    
www.tracoe.com

Managing Director:  
Dr. Christian Zischek

Magistrate's Court, HRB 44548  
Tax No.: 08/658/71878

Your Reference

Our Reference

Date

24. Januar 2023

## URGENT Field Safety Notification

FSCA 2023-01

### Missing IFU of Tracoe Experc Dilation Set (REF 520) in Tracoe Experc Sets Vario, Twist and Twist Plus

Dear Customer,

Tracoe medical GmbH has issued a voluntary Field Safety Corrective Action by sending this Field Safety Notification to point out the absence of instructions for use (IFU) for "Tracoe experc Dilation Sets" (REF 520) in the products Tracoe Experc Set Vario, Twist and Twist Plus due to a packaging list error.

Products mentioned below with the manufacturing date (MAN) 27.12.2021 to 12.01.2023 may be affected.

Our records indicate that you have received some of the affected products mentioned below, whose individual batch numbers are listed on the Confirmation Form attached to this notification.

Tracoe Experc Set Vario with REF numbers:	Tracoe Experc Set Twist with REF numbers:	Tracoe Experc Set Twist Plus with REF numbers:
• REF 420-07 / -08 / -09	• REF 320-07 / -08 / -09	• REF 330-07 / -08 / -09
• REF 421-07 / -08 / -09	• REF 321-07 / -08 / -09	• REF 331-07 / -08 / -09 / -10
• REF 422-08 / -09	• REF 322-07 / -08 / -09	• REF 332-07 / -08 / -09 / -10
• REF 423-07 / -08 / -09	• REF 888-322-07 / -08 / -09	• REF 888-332-07 / -08 / -09 / -10
• REF 424-07 / -08 / -09		
• REF 425-08 / -09		

Mitglied im



Commerzbank Hanau  
IBAN DE63 5064 0015 0240 0604 00  
BIC COBADEFF

Frankfurter Volksbank  
IBAN DE48 5019 0000 6100 0113 00  
BIC FFVBDEFF

Deutsche Bank Frankfurt  
IBAN DE83 5007 0010 0469 4006 00  
BIC DEUTDEFF

## Description of the problem and immediate action required

Tracoe medical GmbH initiated this Field Safety Notice when it was discovered that the IFU of the Tracoe Experc Dilation Set with REF 520 was missing in Tracoe Experc Sets for Vario, Twist and Twist Plus. The IFU has been missing since December 2021.

As of the manufacturing date (MAN) 12.01.2023 the IFU is included.

In order to provide you with all necessary product information contained in the IFU of the Tracoe Experc Dilation Set as quickly as possible, we provide the following options:

1. **Download** the IFU Tracoe Experc Dilation Set, REF 520 from the website.

Please use the web links listed here to access the document directly on our website and print out the necessary information in your national language.

[IFU TRACOE Experc Dilation Set REF 520 \(Atos Medical website\)](#)

[IFU TRACOE Experc Dilation Set REF 520 \(Tracoe Medical website\)](#)

2. **Provision of the booklet** of the IFU Tracoe Experc Dilation Set, REF 520

In case you are not able to download the IFU from the website, we will send you the IFU REF 520 in printed form, which you can add to the existing product package. Please let us know with the Confirmation Form how many printed IFUs you require.

## Forwarding of this Field Safety Notification

This notification should be forwarded to all persons who need to be informed within your institution or another company to which the products affected have been sold.

If applicable, include end users, doctors, nurses, risk managers, supply chain/distribution centres, etc. in the distribution list for this Field Safety Notice.

If you are not a direct end-user, please ask your end-users to complete the confirmation and return it to you.

Please confirm receipt of this Safety Notification by returning the confirmation form attached by e-mail or fax no later than **24. February 2023**.

## Contact

Should you require further information or assistance regarding this measure, please contact

### Customer Service of Tracoe medical GmbH:

**Phone (DE):** +49-6136-9169-145

**Fax:** +49-6136-9169-200

**E-mail:** FSCA@tracoe.com

Tracoe medical GmbH is committed to providing high quality, safe and effective products.

We expressly apologise for any inconvenience this may cause you.

If you have any further questions, please contact your local sales representative or customer service.

Yours sincerely,



Dr. Christian Zischek

Managing Director Tracoe medical GmbH

# Confirmation Form

## Field Safety Notification FSCA 2023-01

**Tracoe experc Set Vario, Tracoe experc Set Twist and Tracoe experc Set Twist Plus  
with missing IFU of REF 520**

**Name of the customer.**      <Kd Name>  
  
   <Kd Straße> <Hausnr.>  
  
   <PLZ> <Ort>

**Customer number:**            <Kd Nummer>

Please fill out this form and send it by fax to +49-6136-9169-200 or by e-mail to FSCA@tracoe.com

**Order of IFU booklets**    [ ] Yes    [ ] No    **Please, enter number of booklets:** \_\_\_\_\_

<p><b><u>For Medical Facility:</u></b></p> <p><input type="checkbox"/> We acknowledge receipt of this Field Safety Notification.</p> <p><b><u>For Distributor:</u></b></p> <p><input type="checkbox"/> We acknowledge receipt of this Field Safety Notification and have forwarded the Field Safety Notification to our customers.</p>
---

Name, First name: \_\_\_\_\_

E-mail: \_\_\_\_\_

Date, Signature: \_\_\_\_\_

**List of affected products:**

REF	LOTs