

VYGON - Prager Ring 100 - 52070 Aachen

Customer

Director Quality Management and Regulatory Affairs

> Tel.: 0241 / 91 30 0 Fax.: 0241 / 91 30 106

Aachen, 20.12.2019

Urgent Recall on bipacingball, Code 1144.05, Batches 051119GD, 141119GD and 221119GD

Dear Madam, dear Sir,

We supply a double sterile packaged product. We have identified a non systematic manufacturing problem, which might lead to potential unsterility at the product's outer pouch. As the inner blister packaging is undamaged, sterility of the product itself is not in danger.

The following products are involved in this recall:

Product Description	Code No.	Batch No.	
bipacingball	1144.05	051119GD	
		141119GD	
		221119GD	

We therefore ask you to immediately withdraw these batches from services and to return them according to our instructions below.

Please acknowledge receipt of this letter and complete and return the attached form by indicating the quantities withdrawn from your institution.

The Federal Institute for Drugs and Medical Devices (BfArM) has been informed about this Field Safety Notice, FSN.

If you should you require further information, please contact your Vygon sales contact, or sales administration via telephone number detailed below :

We apologize for any inconvenience this Recall may cause.

Yours sincerely,



Urgent Recall on bipacingball, Code 1144.05, Batches 051119GD, 141119GD and 221119GD

Please complete and return this form: by fax: xxxxxxxxxx or e-mail to : xxxxxxxxxxx			
Name and address of the institution :			
Full name of the person to contact:			
Function :			
Phone Number:			
ீ Email:			
	I and that the information contained in this field safety / end users of above products within our organization.		
e have the following products available on s	TOCK:		
	If you have units, number of units removed:		
de have the following products available on so			

BIC Code: AACSDE 33