

ASID BONZ GmbH, Hertzstraße 3, 71083 Herrenberg, Germany

Address Customer

ASID BONZ GmbH
Hertzstrasse 3
71083 Herrenberg, Germany

Phone: +49 7032 9562-0
Fax: +49 7032 9562-58
Mail: info@asid-bonz.de
Web: www.asid-bonz.de

12.02.2021 / Kunde / ADM

Safety Information

Dear customer,

with this letter we would like to inform you about the preventive measures taken.

REF: 643020 Urosid® Basic Silicon, two lumen, CH 20, 40 cm long

Based on feedback from the market, that the urinary catheter Urosid® Basic Silicone is not checked for correct filling process including leakage test before use. It is therefore an incorrect application of the product.

We have therefore decided to revise the IFU (to avoid possible complications) and to describe the correct filling process, including the leak test, in more detail.

The clinic advisor responsible for you will be happy to answer any further questions or give you instructions.

Please ensure in your organization, that all users of the above product and other persons to be informed, become aware of this customer information. If you have given the products to third parties, please forward a copy of this information to the third party concerned.

Best regards
ASID BONZ GmbH



Frank Bündler
Head of QM/RA

12.02.2021, safety Information 643020 Urosid® Basic Silicon, two lumen, CH 20, 40 cm long
Customer/ Clinic Advisor/ Customer No.

Please send back to: **ASID BONZ GmbH Fax: 07032 / 9562-58**

We hereby confirm that we have received the customer information on the biliary drainage bag REF: 690630 and all users of the above Products and other persons to be informed have been informed of this customer information.

Place, Date

Signature