

Siemens Healthcare Diagnostics GmbH, SHS EMEA CEECA QT, Siemensstrasse 90,
1210 Vienna

Name	Florian Nemestothy, BSc
Department	SHS EMEA CEECA QT
Mobile	+43 (664) 8855 8770
E-mail	florian.nemestothy@siemens-healthineers.com
Date	April 25, 2025
Document Ref#	PPC25-02.A.OUS

Urgent Field Safety Notice:

BN II System

BN ProSpec System

Atellica NEPH 630 System

Atellica CH Analyzer

Atellica CI Analyzer

N Latex FLC LAMBDA lot 473298/473298A – Incorrect Shelf-Life Information

To whom it may concern,

Siemens Healthineers has confirmed through internal investigation that N Latex FLC LAMBDA lot 473298 and lot 473298A have an incorrect expiry date printed on the vial label, box label and Siemens Certificate of Analysis (CoA).

It was shown that the shelf-life information is erroneously labeled for 12 months longer than the correct expiry date.

Incorrect expiry date: 07-November 2026

Correct expiry date: 07-November-2025

No other lots of N Latex FLC Lambda currently on the market are affected by this issue.

Assay	Siemens Material Number/Unique Device Identification	Lot Number	Manufacturing Date	Expiration Date
N Latex FLC LAMBDA	10482438/ 00842768031427	473298	08-May-2024	07-Nov-2025
		473298A		

Impact to Results

If the affected kit lots are used beyond the correct expiry date, QC recovery will indicate the issue, potentially leading to an apparent delay in testing while resolving a QC failure.

Customer Actions:

Perform the instructions provided below:

- The affected N Latex FLC LAMBDA lots 473298 and 473298A must not be used after 2025-11-07.

Please review this letter with your Medical Director.

- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- Please retain this letter with your laboratory records and forward this letter to those who may have received this product.


Please Note:

- The Siemens CoA has been revised and a correct version is available in the Document Library and attached to this letter.
- For the Atellica NEPH 630 System the expiration date has being corrected on the new Lot Data version from 2025-04.
- For the BN II and BN ProSpec Systems no expiration dates are being displayed within the software.
- For the Atellica CH and Atellica CI analyzer, the reagent must be transferred into Atellica CH EMPTY P1 and manually identified, therefore please use the correct expiration date (2025-11-07).

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Sincerely yours,

Siemens Healthcare Diagnostics GmbH



Electronically signed by: Martina
Schreiber
Reason: I am approving this document
Date: Apr 26, 2025 17:15 GMT+2

ppa. Mag. Martina Schreiber



Electronically signed by: Michael
Edetsberger
Reason: I have reviewed this document
Date: Apr 25, 2025 20:55 GMT+2

i.V. Dr. Michael Edetsberger